

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Name and address of the applicant

Control units

EQON AS
Veritasveien 25
4007 Stavanger
NORWAY

Name and address of the manufacturer

EQON AS
Veritasveien 25, 4007 Stavanger, NORWAY

Name and address of the factory

EQON AS
Veritasveien 25, 4007 Stavanger, NORWAY

Ratings and principal characteristics

Power Supply Controller:	24 VDC
Power Supply Main:	230 VAC +/- 10 %, 50/60 Hz
Maximum Power Output:	32 A / 230 VAC
Protection class:	I

Model/type Ref.

EQON ONE

A sample of the product was tested and found to be in conformity with

IEC 61010-1:2010
IEC 61010-1:2010/AMD1:2016

as shown in the Test Report Ref. No. which forms part of this certificate

713282038

Page 1 of 2

This CB Test Certificate is issued by the National Certification Body

CB 121120 0001 Rev. 00

Date, 2023-05-15



(Xiaoxin Liang)

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany



Product Service

Ratings and principal characteristics (continued)

Remarks/CoA (e.g. nomenclature; conditions of acceptability):

Condition of Acceptability:

- This equipment is for indoor use in non-hazardous locations, operated by qualified personnel skilled in its use.
- When installing requirements of test standards and installation guide must be fulfilled
- Factory wiring only
- Must be supplied by Recognized/Certified Power Supply with fulfil the requirements for RI/DI Isolation between SELV and Primary
- An External Fuse is required for the 230 VAC Supply

Page 2 of 2

This CB Test Certificate is issued by the National Certification Body

CB 121120 0001 Rev. 00

Date, 2023-05-15



(Xiaoxin Liang)